



## DECLARATION OF CONFORMITY

For the following equipment:  
Electric Powered Adjustable Patient Lifters

Model :FLP500,FLNP500,PR51FLP600,FLNP600

Is herewith confirmed to comply with the requirements set out in the council Directive on the harmonization of the laws of UK MDR2002.

For the evaluation regarding the **Class I** product safety aspects, the following standards were applied :  
ISO13485:2016, ISO10535:2006, ISO14971:2019, IEC60601-1:2014  
EN 60601-1-2:2015

The following company is responsible representative for this declaration:

Advena Ltd

Pure Offices, Plato Close, Tachbrook Park, Warwick, CV34 6WE. UK

Person responsible for making this declaration:

**APEX HEALTH CARE MFG., INC.**

(Manufacturer Name)

**No.6, Industrial 2nd Road, Tou Chiao Industrial Park, Min Hsiung Shiang, Chia-YiHsien, Taiwan,R.O.C.**

(Place, Manufacturer Address)

John Cheng/ Director

June 21, 2023.

(