

Test Report SL52025233239101TX Date: April 13,2020 Page 1 of 2 JIANGSU NEWVALUE MEDICAL PRODUCTS CO.,LTD

FLOOR THIRD BUILDING AUXILIARY, FLOORS 1-3 BUILDING MAIN G35 TAIZHOU CHINA MEDICAL CITY

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description (A) Mask(It consists mask body, mask belt and nose clip, The mask body has

three layers: inner layer (spunbond layer), middle layer (melt spraying layer),

outer layer (spunbond layer), nose clip made of bendable materials)

Composition : Non-woven Sample Color : (A)Blue

Proposed Care Instruction:

Sample Receiving Date : Mar 23,2020

Testing Period : Mar 26,2020 - Apr 13,2020

Test Result(s) : For further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant

Signed for and on behalf of SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx.and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <a href="http://www.sgs.com/en/Terms-and-Conditions/Terms-and-

3"Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国 • 上海 • 徐汇区宜山路889号3号楼 邮編: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com



Test Report Test Result

SL52025233239101TX

Date: April 13,2020

Page 2 of 2

Medical Face Masks-Requirements and Test Methods

(EN 14683:2019)

Clause 5.2.2 Bacterial filtration efficiency (BFE)"

	1#	2#	3#	4#	5#
(BFE), %	100	100	99.9	100	99.9

Remark: Performance Requirement: Type I>95%, Type II>98%, Type IIR >98% **: The test was carried out by external laboratory assessed as competent

Clause 5.2.3 Breathability

(EN 14683 :2019 Annex C, Flow rate 8 I/min)

Sample: A

	1#	2#	3#	4#	5#
Differential pressure △P (Pa/cm²)	36.7	33.1	34.3	32.9	29.2

Remark: Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²

Clause 5.2.5 Microbial Cleanliness

(EN 14683: 2019 Annex D)

5# 1# 2# 3# 4# CFU/a <1 <1

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx.and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at https://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx.attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's Instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the

3 Building, No. 839, Yishan Road, Xuhui District Shanghai, China 200233 中国・上海・徐汇区宜山路889号3号楼 邮編: 200233

t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763

e sgs.china@sgs.com



Test Report SL52025233239701TX Date: April 10,2020 Page 1 of 2

JIANGSU NEWVALUE MEDICAL PRODUCTS CO.,LTD.

FLOOR 1-3 OF MAIN BUILDING AND FLOOR 3 OF AUXILIARY BUILDING OF BUILDING G35 OF TAIZHOU CHINA PHARMACEUTICAL CITY

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Mask (It consists mask body, mask belt and nose clip, The mask body has

three layers: inner layer (spunbond layer), middle layer (melt spraying layer),

outer layer (spunbond layer), nose clip made of bendable materials)

Composition : (A)nonwoven fabric

Sample Color : (A)Blue

Proposed Care Instruction : -

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : Mar 23,2020

Testing Period : Mar 26,2020 - Apr 10,2020

Test Result(s) : For further details, please refer to the following page(s).

Signed for and on behalf of SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.ags.com/en/Terms-and-Conditions.aspx and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at http://www.ags.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not excertate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. (86-755) 8307 1443.

3"Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666 f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com



Test Report

SL52025233239701TX

Date: April 10,2020

Page 2 of 2

Test Result

Medical Face Masks-Requirements and Test Methods

(EN 14683:2019)

Clause 5.2.4 Splash Resistance

Penetration on inside surface

(EN 14683.20	119)							
	Splash Resis 2004, Pressur							
Sample: A								
Penetration	on inside sur	face						NO
1#	2#	3#	4#	5#	6#	7#	8#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	
9#	10#	11#	12#	13#	14#	15#	16#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	
17#	18#	19#	20#	21#	22#	23#	24#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	
25#	26#	27#	28#	29#	30#	31#	32#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	

Number of Pass:

Overall result: Acceptable

Remark:

- Performance Requirement Type I: N/A, Type II: N/A, Type IIR:≥16.0kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is 300±10mm.

32

- 3) Condition and Test temperature (21±5)° C, relative humidity (85±10)%
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results

The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report

